



FOR IMMEDIATE RELEASE

**HEALTH SCIENCES AUTHORITY
PRESS RELEASE
22 NOVEMBER 2006**

HSA ALERTS PUBLIC ON RECALL OF SPECIFIC LOTS OF AMO COMPLETE® BRAND MULTI-PURPOSE SOLUTIONS SOLD IN SINGAPORE

The Health Sciences Authority (HSA) has been notified by Advanced Medical Optics (AMO), Singapore, on 21 November 2006, on the voluntary market recall of three specific lots of its multi-purpose solutions in Singapore.

Product Description	Lot No.
AMO Complete Protec™ Multipurpose Solution (350ml)	ZB02704
AMO Complete Protec™ Multipurpose Solution (350ml)	ZB03247
AMO Complete MoisturePLUS™ Multipurpose Solution (60ml)	ZB03194

2 The three affected lots are being recalled in Singapore by AMO as a precautionary measure because of their concern over the sterility status of the product. Information available at this point of time indicates that the problem lies with AMO's manufacturing plant in China that manufactured these lots. For further information on the recalled lots, please refer to the press release by AMO on the company's website at www.amo-inc.com.

3 HSA has directed the company to confirm the sterility status of other lots of these products currently available in Singapore. HSA is also coordinating its regulatory follow-up with its counterpart agencies including the US FDA.

4 Consumers who have these recalled lots should discontinue use immediately and contact AMO's hotline at 6844 8833 for further queries. Those who experience symptoms of an eye infection such as redness, pain, tearing, increased light sensitivity, blurry vision, discharge or swelling are advised to consult their eye care provider or an ophthalmologist immediately.

5 To date HSA has not received any reports of adverse events associated with the recalled product lots in Singapore. HSA will continue to closely monitor the situation, and will alert the public and take any appropriate action to protect public health and safety.

**HEALTH SCIENCES AUTHORITY
22 NOV 2006**