Page 1 of 7



### **URGENT: MEDICAL DEVICE RECALL (REMOVAL)**

VICRYL RAPIDE™ (Polyglactin 910) SUTURE (SPECIFIC CODES AND LOTS LISTED BELOW)

September 28, 2017

Dear Customer, Operating Room Supervisors, Materials Management Personnel, and Chief of Surgery:

## PLEASE DISTRIBUTE THIS INFORMATION TO ALL STAFF WITHIN YOUR FACILITY WHO USE VICRYL RAPIDE™ (Polyglactin 910) SUTURE

At Ethicon, Inc. ("Ethicon"), our first priority is to our customers and their patients, and that includes the safe and effective use of our products.

Ethicon has initiated a medical device recall (removal) of <u>certain lots</u> of VICRYL RAPIDE™ (Polyglactin 910) Suture. We have identified that certain lots of sizes 2-0 and 4-0 for VICRYL RAPIDE™ SUTURE (Product Codes and Lot Numbers listed in table below) exhibited suture damage due to a manufacturing equipment issue. The manufacturing equipment issue has been corrected.

The medical assessment concluded that this situation could result in a potential superficial wound dehiscence or contribute to impaired wound healing but only occasionally and/or under unusual circumstances.

Ethicon has not received any reports of Adverse Events associated with the product subject to this recall (removal).

Health care practitioners that have treated patients using VICRYL RAPIDE™ SUTURE from the lots listed in the table below, should continue to follow those patients post-operatively in the usual manner.

This medical device recall (removal) has been communicated to the U.S. Food and Drug Administration (FDA).





Page 2 of 7

## EFFECTIVE IMMEDIATELY – DO NOT USE OR DISTRIBUTE THE FOLLOWING PRODUCT CODES/LOTS:

PRODUCT NAME	PRODUCT CODE	PRODUCT LOT	DESCRIPTION/SIZE		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	MPVR4260H	AH6847	Size 4-0 VICRYL RAPIDE Undyed 75cm PS-2 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	V2140H	AH6862	Size 4-0 VICRYL RAPIDE Undyed 70cm RB-1 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	V2140H	AH7675	Size 4-0 VICRYL RAPIDE Undyed 70cm RB-1 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	V2140H	AH9170	Size 4-0 VICRYL RAPIDE Undyed 70cm RB-1 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR2217	AH7884	Size 2-0 VICRYL RAPIDE Undyed 90cm V-34 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR2217	AH9680	Size 2-0 VICRYL RAPIDE Undyed 90cm V-34 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR2253	AH8749	Size 2-0 VICRYL RAPIDE Undyed 75cm FS-1 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR2253	AH9686	Size 2-0 VICRYL RAPIDE Undyed 75cm FS-1 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR2253	AJ0204	Size 2-0 VICRYL RAPIDE Undyed 75cm FS-1 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR2253	AH9968	Size 2-0 VICRYL RAPIDE Undyed 75cm FS-1 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR2253	AJ1166	Size 2-0 VICRYL RAPIDE Undyed 75cm FS-1 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR2279	AH7676	Size 4-0 VICRYL RAPIDE Undyed 75cm V-4 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR2294	AH7679	Size 4-0 VICRYL RAPIDE Undyed 75cm FS-3 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR2294	AH7684	Size 4-0 VICRYL RAPIDE Undyed 75cm FS-3 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR2297	AH7680	Size 4-0 VICRYL RAPIDE Undyed 75cm FS-2 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR2297	AH7711	Size 4-0 VICRYL RAPIDE Undyed 75cm FS-2 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR2297	AH7686	Size 4-0 VICRYL RAPIDE Undyed 75cm FS-2 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR2297	AJ0770	Size 4-0 VICRYL RAPIDE Undyed 75cm FS-2 Needle		
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR2297	AJ0794	Size 4-0 VICRYL RAPIDE Undyed 75cm FS-2 Needle		



Johnson & Johnson Medical Singapore a division of Johnson & Johnson Pte Ltd.

2 Science Park Drive, #07-13, Ascent, Singapore Science Park 1, Singapore 118222

Tel: +65 6918 8000



PRODUCT NAME	PRODUCT CODE	PRODUCT LOT	DESCRIPTION/SIZE
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR496	AH7694	Size 4-0 VICRYL RAPIDE Undyed 18" PS-2 Needle
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VR945	AH6611	Size 2-0 VICRYL RAPIDE Undyed 36" CT-1 Needle
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VRI253H	AH8642	Size 2-0 VICRYL RAPIDE Undyed 75cm FS-1 Needle
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	VRI253H	AH8644	Size 2-0 VICRYL RAPIDE Undyed 75cm FS-1 Needle
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	W9941	AH8379	Size 2-0 VICRYL RAPIDE Undyed 75cm FS Needle
VICRYL RAPIDE™ (Polyglactin 910) SUTURE	W9941	AH9666	Size 2-0 VICRYL RAPIDE Undyed 75cm FS Needle

#### **IDENTIFICATION OF PRODUCT SUBJECT TO THIS RECALL (REMOVAL):**

Product subject to the recall (removal) in your inventory can be identified by product code and lot number (see product code and lot listing above). All unused VICRYL RAPIDE™ SUTURE products subject to this recall (removal) are required to be returned. The product code and lot number can be determined by using the Product Identification Tool within Attachment 1.

#### **ACTION REQUIRED:**

- 1. Examine your inventory immediately to determine if you have product subject to this recall (removal) on hand and quarantine such product(s).
- 2. Remove the product subject to this recall (removal) and communicate the issue to relevant operating room or materials management personnel, or anyone else in your facility who needs to be informed.
- 3. If any product subject to this recall (removal) has been forwarded to another facility, contact that facility to arrange return.
- 4. Complete the Business Reply Form (BRF) (Attachment 2) confirming receipt of this notice and return it to your local Ethicon Sales Representative within three (3) business days. Please return the BRF even if you do not have product subject to this recall (removal).
- 5. Keep this notice visibly posted for awareness until all product subject to this recall (removal) has been returned to Johnson & Johnson Medical Singapore. While processing your returns, please maintain a copy of this notice with the product subject to this recall (removal) and keep a copy for your records.



Johnson & Johnson Medical Singapore a division of Johnson & Johnson Pte Ltd.

2 Science Park Drive, #07-13, Ascent, Singapore Science Park 1, Singapore 118222

Tel: +65 6918 8000





6. Customers are required to return all unused VICRYL RAPIDE™ SUTURE products that are in their inventory immediately. Only unused VICRYL RAPIDE™ SUTURE product subject to this recall (removal) returned by <u>December 15, 2017</u> will be eligible for replacement. Any product subject to this recall (removal) returned after <u>December 15, 2017</u> will not be eligible for replacement.

If you require any assistance with returning product, please contact your local Ethicon Sales Representative.

We recognize the recall (removal) of the VICRYL RAPIDE™ SUTURE may be disruptive to your facility and we apologize for any inconvenience this may cause.

If you have additional questions regarding this recall (removal) or to report any customer complaints, please contact your local Ethicon Sales Representative.

Healthcare professionals are advised to report any adverse events and/or suspected adverse reactions or quality problems associated with the use of this product to your Ethicon Sales Representative or contact Goh Jun Lee at Tel: +65 6918 7781 immediately or email to <a href="https://doi.org/10.2006/journal.org/linearing-nc-nd-1.00">DL-JNJSG-One-JnJ-SG-Quality@ITS.JNJ.com</a>.

Alternatively, you may also report any adverse events to the Vigilance and Compliance Branch, Health Products Regulation Group at Tel: 6866 3538, Fax: 6478 9069, or report online via www.hsa.gov.sg/ae\_online. Events that are reported to Johnson & Johnson Pte Ltd will be investigated and subsequently reported to HSA.

#### **Attachments:**

Attachment 1: Product Identification Tool Attachment 2: Business Reply Form

Yours sincerely,



Lee Ching Hwee
Manager, Regulatory Affairs

cc: Chairman Medical Board
Relevant Head of Departments



Page 5 of 7




# **ATTACHMENT 1:** Product Identification Tool for VICRYL RAPIDE™ SUTURE (SPECIFIC CODES AND LOTS LISTED IN TABLE ABOVE)

This tool will help customers identify the product subject to this recall (removal) by using the package labels. This document applies to the VICRYL RAPIDE™ SUTURE, sales unit carton, and foil pouch for Product Code VR2297.

**Please note:** The labels listed below are used for illustration purposes only. <u>Refer to the table listed above</u> for a list of product codes and lot numbers subject to this recall (removal).

#### VICRYL RAPIDE™ SUTURE SALES UNIT CARTON (CONTAINING 36 EACHES)

#### FRONT of SALES UNIT CARTON (PRODUCT CODE VR2297)





Tel: +65 6918 8000





# **ATTACHMENT 1:** Product Identification Tool for VICRYL RAPIDE™ SUTURE (SPECIFIC CODES AND LOTS LISTED IN TABLE ABOVE)

#### VICRYL RAPIDE™ SUTURE FOIL POUCH (CONTAINING 1 EACH)

#### FRONT of FOIL POUCH (PRODUCT CODE VR2297)





2 Science Park Drive, #07-13, Ascent, Singapore Science Park 1, Singapore 118222

Tel: +65 6918 8000

Page 7 of 7



### **ATTACHMENT 2:** Business Reply Form (BRF)

Product Inventory: - please check one

Your timely response to this customer notification is requested. Please complete and submit this form to your local Ethicon Sales representative or email it to RA-ITSUS-JJMSingapo@ITS.JNJ.COM within 3 business days, even if you do not have product subject to this recall (removal) to return.

If you have product subject to this recall (removal) to return, please indicate the product code, lot number and quantity of each being returned in the table below and make a photocopy of your completed Business Reply Form and enclose with your return - Please refer to the above listed table of product subject to this recall (removal). Thank you for your cooperation.

Device Name	Product Code	Lot Nu	ımber	Quantity Return (in "Eaches"	
VICRYL RAPIDE™ (Polyglactin 910) SUTURE					
lame and Title of Person Comp	leting Business Reply	Form:	Telephon	e Number:	
acility Name & Company Stam	p:		Date:		
igned*:					



\_\_\_\_\_