

Urgent Field Safety Notice (Field Safety Corrective Action)
CORAIL® AMT Neck Trials Surgical Instruments

<u>Product Name</u>	<u>Product Code</u>
CORAIL AMT NECK SEG 125D STD	L94003
CORAIL AMT NECK SEG 125D KLA	L94004
CORAIL AMT NECK SEG 135D STD	L94005
CORAIL AMT NECK SEG 135D KHO	L94006
CORAIL AMT NECK SEG 135D SHORT	L94007
CORAIL AMT STANDARD NECK SEGMENT	L20431
CORAIL AMT COXA VARA NECK SEGMENT	L20432
CORAIL AMT HIGH OFFSET NECK SEGMENT	L20433

FSCA-identifier: PIE-1125109 Version 2

Type of Action: Field Safety Notice (Field Safety Corrective Action)

Date: 31 Jul 2018

Attention: Trust Chief Executives, the Clinical Director of the Orthopaedic Department, the Orthopaedic Theatre Manager, the Safety Liaison Officer, General Managers of Private Sector Hospitals, Distributors

DePuy France, SAS is issuing a Field Safety Notice/Field Safety Corrective Action for the above-mentioned product codes. This correction is being issued to address the potential for debris/material to be found behind the O-Rings for some CORAIL Neck Trials.

This is an update to the Field Safety Notice/ Field Safety Corrective Action issued in May 2018 to include additional product codes.

A company representative will contact you shortly to rework affected devices at your facility or provide directions on replacing affected devices with reworked devices. Until the company reworks the affected units



Figure 1: Image of Corail Neck Trial

and to reduce the possibility of debris/material being left behind the O-Rings, the company recommends adhering to the instructions for use in IFU-W90946 Rev B.

Cleaning instructions from IFU-W90946 Rev B are included for your reference as Attachment A in this Field Safety Notice.

Type of Device

The below mentioned CORAIL Neck Trials are surgical instruments used in CORAIL total and partial hip arthroplasty. No other CORAIL devices are affected by this Field Safety Notice.

Affected Product

Product Code	Lot Number	GTIN No.	Model Name
L94003	All Lots	10603295325147	CORAIL AMT NECK SEG 125D STD
L94004	All Lots	10603295325154	CORAIL AMT NECK SEG 125D KLA
L94005	All Lots	10603295325161	CORAIL AMT NECK SEG 135D STD
L94006	All Lots	10603295325178	CORAIL AMT NECK SEG 135D KHO
L94007	All Lots	10603295325185	CORAIL AMT NECK SEG 135D SHORT
L20431	See Attachment B	10603295258216	CORAIL AMT STANDARD NECK SEGMENT
L20432	See Attachment B	10603295258223	CORAIL AMT COXA VARA NECK SEGMENT
L20433	See Attachment B	10603295258230	CORAIL AMT HIGH OFFSET NECK SEGMENT

Clinical Implications and Patient impact

Fifteen complaints have been received related to debris behind the O-Ring. The company evaluated this issue and to date determined that none of these complaints have resulted in patient harm and found no increased risk to the patient.

Field Safety Corrective Action

As a precautionary measure, the company determined that reworking the affected devices and removing the O-Rings is the appropriate corrective action. Any unit of the affected product codes containing an O-Ring will have the O-Ring removed by DePuy representative. The function of the device is unaffected by this change as DePuy currently markets products of the same design (i.e. without an O-Ring). This Field Safety Notice provides instructions for notifying medical facilities that may have used, purchased, or received the affected units. The purpose of this Field Safety

Corrective Action is to alert medical facilities of the rework plan to remove the O-Ring from the affected devices.

Please undertake the following urgent actions:

- Please continue to follow the instructions for use in IFU-W90946 Rev B regarding cleaning of these devices.
- Medical facilities are to determine if any of the affected instruments are on hand, and contact their Sales Consultant to arrange for rework or replacement of these instruments.
- Review this notice and complete the Acknowledgement section (Attachment B) to signify that your facility has been informed of this Field Safety Notice. Return the completed Acknowledgement to your Sales Consultant within five (5) working days of this notice.
- Retain a copy of the completed Acknowledgement Form in your files along with this notice.
- Notify surgeons at your facility by providing them with a copy of this notice to ensure surgeons are aware of this Field Safety Notice.
- Share this notice with others in your facility who need to be informed.
- If any affected product has been forwarded to another facility, contact that facility immediately to communicate this field action with the facility/facilities. Inform your Sales Consultant if further facilities are affected.

Transmission of this Field Safety Notice:

This notice has been sent to you because our records indicate that you have received the affected product.

This notice needs to be passed on to all those who need to be aware within your organization.

For any enquiries regarding this Field Safety Notice, contact your Sales Consultant.

Notification of this FSN has been provided to the appropriate Regulatory Agency.

Yours sincerely,



Lee Ching Hwee

Manager, Regulatory Affairs

cc: Chairman Medical Board

Relevant Head of Departments

Johnson & Johnson
MEDICAL

Johnson & Johnson Medical Singapore
a division of Johnson & Johnson Pte Ltd.
2 Science Park Drive, #07-13, Ascent, Singapore Science Park 1, Singapore 118222
Tel: +65 6918 8000
Business Reg No. 52836279L Company Reg No. 197402104W

ATTACHMENT A

Extract from IFU-W90946 Rev B:

From Page 6 Section G of IFU-W90946 Rev B - Manual Cleaning: All Devices:

- Prepare an enzymatic cleaning solution in accordance to the manufacturer's instructions.
- Soak soiled devices for a minimum recommended time specified by the enzymatic cleaning solution manufacturer or 5 minutes, whichever is longer.
- Prepare a pH neutral (pH 7-9) detergent cleaning solution in accordance to the manufacturer's instructions.
- Use a soft non-metallic bristle brush (plastic bristles, like nylon) to thoroughly scrub all traces of blood and debris from the device surfaces for one minute.
- Rinse the device with warm, 30°C - 40°C (85°F – 104°F), tap water for a minimum of one minute and until visual evidence of debris, soil, and cleaning solution are gone.
- Pay particular attention to thoroughly flush lumens, articulating areas, and flexible segments with warm, 30°C - 40°C (85°F – 104°F), tap water.
- Ultrasonically clean the device components for 10 minutes in neutral pH detergent (pH 7-9), prepared in accordance with the manufacturer's instructions.
 - NOTE: Ultrasonic cleaning is only effective if the surface to be cleaned is immersed in the cleaning solution. Air pockets will decrease the efficacy of ultrasonic cleaning. Be sure to minimize air pocket or bubble formation by flushing lumens, cavities, crevices or springs with cleaning solution while the instrument is immersed in the ultrasonic cleaner tank.
- Rinse the device components with warm, 30°C - 40°C (85°F – 104°F), tap water for a minimum of one minute and until visual evidence of debris, soil, and cleaning solution are gone.
- Perform a final rinse with Reverse Osmosis Deionized (RODI) or Purified (PUR) Water.
- Dry the device components immediately after final rinse with a clean towel or clean compressed air until visibly dry.

From Page 8 Section L of IFU-W90946 Rev B - Automated Cleaning:

- Prepare an enzymatic cleaning solution in accordance to the manufacturer's instructions.
- Soak devices for a minimum recommended time specified by the enzymatic cleaning solution manufacturer or 5 minutes, whichever is longer.
- All devices should be pre-cleaned in accordance with the appropriate Manual Cleaning Instructions section.

- Rinse the device with warm, 30°C - 40°C (85°F – 104°F), tap water for a minimum of one minute and until visual evidence of debris, soil, and cleaning solution are gone.
- Pay particular attention to thoroughly flush lumens, articulating areas, and flexible segments with warm, 30°C - 40°C (85°F – 104°F), tap water for a minimum of one minute and until visual evidence of debris, soil, and cleaning solution are gone.
- Load the device components so that the lumens can drain.
- Clean, using the “INSTRUMENTS” cycle in a validated washer disinfector and a pH neutral cleaning agent intended for use in automated cleaning using the minimum cycle parameter set below:

Phase	Time (Minutes)	Temperature	Detergent Type
Pre Wash	2:00	Cold Tap Water	N/A
Enzyme wash	1:00	< 40°C	Enzymatic Cleaner
Wash	2:00	66°C	Neutral pH Detergent
Rinse	0:15	> 40°C	N/A
Thermal Decontamination*	5:00	> 93°C	N/A
Dry	7:00	115.5°C	N/A

* Reverse Osmosis Deionized (RODI) or Purified (PUR) Water

Attachment B

This attachment lists the lots numbers of product codes L20431, L20432 and L20433 manufactured with an O-Ring present.

Product Code	Lot	Product Code	Lot	Product Code	Lot	Product Code	Lot
L20431	1212901	L20431	2163174	L20431	2490514	L20431	2774410
	1218481		2249115		2513465		2795717
	1226220		2249116		2513466		2795718
	1230761		2252364		2526363		2795719
	1230762		2274452		2526364		2811227
	1700575		2275487		2539884		2811228
	1781381		2284306		2539885		2811233
	1812714		2284307		2565583		2811234
	1817780		2309644		2572825		2836707
	1860264		2309645		2572826		2836708
	1865892		2309646		2602513		2836709
	1874689		2336389		2602514		5001519
	1874692		2336390		2605764		5001520
	1874695		2362086		2605765		5001521
	1885306		2362087		2643408		5001522
	1899212		2380243		2643409		5005846
	1899213		2384473		2648066		5005847
	1910187		2384474		2648067		5008238
	1979239		2386339		2673433		5008239
	2002010		2401664		2673434		5009208
	2024019		2401665		2691497		5016192
	2043428		2416615		2691501		5016193
	2078239		2416617		2739089		5016194
	2102693		2416618		2739090		5016195
	2129111		2465273		2749282		5016196
	2153317		2465274		2749283		5019838
	2153318		2465275		2774407		5019841
	2153319		2465287		2774408		1812714A
	2153320		2490513		2774409	L20432	1212912

Product Code	Lot	Product Code	Lot	Product Code	Lot	Product Code	Lot
L20432	1216680	L20432	2284305	L20432	2643412	L20432	5012220
	1230766		2309647		2673435		5016191
	1700576		2309648		2673436		5017033
	1812715		2309649		2673437		5017035
	1812716		2336392		2691532		1812716A
	1817781		2336393		2691533	L20433	1216681
	1860265		2384475		2739091		1216682
	1865893		2384476		2739092		1230767
	1874690		2386340		2749286		1230768
	1874693		2401666		2774417		1230769
	1874696		2416616		2774418		1700574
	1885307		2416619		2774419		1700577
	1899214		2416620		2795761		1781383
	1899215		2465280		2795762		1812717
	1910188		2465288		2795763		1817782
	1979242		2481312		2811229		1860266
	2002011		2490516		2811230		1865894
	2024020		2511420		2811235		1874691
	2043429		2526365		2811236		1874694
	2078240		2565598		2836712		1874697
	2102701		2565705		2836713		1885308
	2129112		2565706		2836714		1899216
	2153321		2572839		5002114		1899217
	2153322		2572840		5002115		1910189
	2153323		2605839		5002116		2002012
	2153324		2605840		5005843		2024021
	2173202		2609135		5005844		2043430
	2284268		2609141		5005845		2065508
	2284304		2643411		5012219		2078241

Product Code	Lot	Product Code	Lot	Product Code	Lot
L20433	2102703	L20433	2465284	L20433	2795714
	2102704		2465285		2795715
	2129113		2465289		2795716
	2142423		2487632		2811225
	2153325		2490517		2811226
	2153326		2490518		2811231
	2173203		2513467		2811232
	2249175		2515079		2823044
	2249176		2526367		2836720
	2249177		2526368		2836721
	2252365		2565627		2836722
	2274507		2565714		5002110
	2284309		2565715		5002111
	2284310		2572844		5002112
	2309652		2572845		5002113
	2309653		2572846		5005841
	2309654		2609035		5005842
	2336394		2609037		5008236
	2336395		2617683		5008237
	2362088		2617684		5009207
	2362089		2643413		5012217
	2380244		2643414		5012218
	2384477		2673438		5016197
	2384478		2673439		5016198
	2401667		2673440		5017028
	2416621		2691548		5017030
	2416622		2691549		5019843
	2424599		2749290		1812717A
	2465283		2761213		

ATTACHMENT C

This Letter acknowledges receipt of the Field Safety Notice related to FSCA Identifier: PIE-1125109 Version 2.

Product Code	Lot Number	Model Name
L94003	All Lots	CORAIL AMT NECK SEG 125D STD
L94004	All Lots	CORAIL AMT NECK SEG 125D KLA
L94005	All Lots	CORAIL AMT NECK SEG 135D STD
L94006	All Lots	CORAIL AMT NECK SEG 135D KHO
L94007	All Lots	CORAIL AMT NECK SEG 135D SHORT
L20431	See Attachment B	CORAIL AMT STANDARD NECK SEGMENT
L20432	See Attachment B	CORAIL AMT COXA VARA NECK SEGMENT
L20433	See Attachment B	CORAIL AMT HIGH OFFSET NECK SEGMENT

(Please check as appropriate)

- ☐ Yes, I have received the FSN
- ☐ No affected products in stock.
- ☐ Yes, I have devices of the affected product code(s).

Please submit a copy of this completed document to your Sales Consultant.

Product Code	Lot Number	Quantity in stock

Name, Sign and Date 	Hospital Name, Address, Phone and Stamp
--	--