

BD Life Sciences  
Diagnostic Systems  
7 Loveton Circle  
Sparks, MD 21152  
tel: 410-316-4000  
fax: 410-316-4066



## URGENT PRODUCT RECALL

June 2016

Product Name	Cat. Number	Unit	Lot Number
BD PhoenixSpec™ Calibrator Kit	440911	each	5364776
BD PhoenixSpec™ Calibrator Kit	440911	each	5327782
BD PhoenixSpec™ AP Calibrator Kit	441951	each	5345796

Dear Microbiology Lab Manager,

BD has determined that the BBL PhoenixSpec Calibrator kits from the above referenced lots are labeled with an incorrect expiration date (year). The expiration date printed on the individual calibration tubes within the kit are correct but difficult to read on a portion of the tubes.

Lot Number	Exp. Date on Kit	Correct Exp. Date
5364776	October 30, 2018	October 30, 2016
5327782	October 30, 2018	October 30, 2016
5345796	October 30, 2018	October 30, 2016



The calibrator kits will perform as intended if utilized to the correct expiration date of October 30, 2016. BD recommends that the PhoenixSpec Calibrator kits from the above referenced lots be discarded by October 30, 2016. If you wish to discard the product now as a precaution, BD will replace the product.


Our records indicate you may have been shipped a portion of the above-referenced product between January 6, 2016 and April 12, 2016.

Please complete the attached form **whether or not you have any inventory remaining** so that we may acknowledge your receipt of this notification. It is not necessary to contact BD by telephone to receive replacements. Simply complete and fax the form to: Attention: Regulatory Compliance, Fax: 410-316-4258.


If further assistance is needed, please contact the BD Customer Service Department at 1-800-638-8656 ext. 3. For all other inquiries, please contact the BD Technical Services Department at 1-800-638-8663 ext. 2.

Please accept our apology for any inconvenience this may cause. BD is committed to providing you with the highest quality products. Thank you for your continued support.

Regards,



Emily Howard  
Manager, Regulatory Compliance



Thierry Guillet  
WW Marketing Manager ID/AST, Life Sciences

Fax or Email to: BD Regulatory Compliance

Fax #: 410-316-4258

Email: BDRC2@bd.com

**PLEASE RETURN THIS FORM SO THAT WE MAY ACKNOWLEDGE YOUR RECEIPT  
OF THIS NOTIFICATION (MC 640)**

Please do not send multiple forms.

Product Name	Cat. Number	Unit	Lot Number	Exp. Date
BD PhoenixSpec™ Calibrator Kit	440911	each	5364776	October 30, 2018
BD PhoenixSpec™ Calibrator Kit	440911	each	5327782	October 30, 2018
BD PhoenixSpec™ AP Calibrator Kit	441951	each	5345796	October 30, 2018

Facility: \_\_\_\_\_

Please use full, current facility name. Do not use initials.

Street Address: \_\_\_\_\_

City: \_\_\_\_\_ State: \_\_\_\_\_ Zip: \_\_\_\_\_

Contact Person: \_\_\_\_\_

Telephone No.: \_\_\_\_\_ Fax No.: \_\_\_\_\_

Total number of affected units on hand:

Product Name	Cat. Number	QTY for Replacement Each- 1 kit
BD PhoenixSpec™ Calibrator Kit	440911	
BD PhoenixSpec™ AP Calibrator Kit	441951	

FOR BD USE ONLY		
BILL TO ACC'T NO. 9000000446	ORDER REASON 517	DELIVERY P.O. NO.
SHIP TO ACC'T NO.	JUSTIFICATION CODE	INVOICE P.O. NO.