

August 03, 2018

To: Risk Managers

Cc: Chairman Medical Board and relevant Head of Departments

Subject: **URGENT MEDICAL DEVICE RECALL – LOT SPECIFIC**

Affected Product: Juggerknot Long Flex Drill Bit with Sleeve Nitinol Stainless Steel

Item Number	Lot Numbers	
110016992	053940	545740
	067490	664610
	165770	676180
	165830	688250
	415590	860050

Zimmer Biomet is conducting a lot specific medical device recall for Juggerknot Long Flex Drill Bit with Sleeve Nitinol Stainless Steel due to the expiration date being incorrectly listed on the label. The labels have a 10 year shelf life but the products have a 5 year expiry period. These devices are still within their 5 year expiry period and are being removed prior to reaching their actual expiration date.

Risks		
Describe immediate health consequences (injuries or illness) that may result from use of or exposure to the product issue.	Most Probable	Highest Severity
	None	None
Describe long range health consequences (injuries or illness) that may result from use of or exposure to the product issue.	Most Probable	Highest Severity
	None	Infection, Surgical Intervention (Biological Response)

Our records indicate that you may have received one or more of the affected products. The affected units were distributed between May 2014 and June 2017.

Risk Manager Responsibilities:

1. Review this notification and ensure that affected personnel are aware of the contents.
2. If you have affected product at your facility, assist your Zimmer Biomet sales representative and quarantine all affected product. Your Zimmer Biomet sales representative will remove the affected product from your facility.
3. Complete **Attachment 1 – Certificate of Acknowledgement** and send to CorporateQuality.PostMarket@zimmerbiomet.com. This form must be returned even if you do not have affected products at your facility.



4. Retain a copy of the acknowledgement form with your recall records in the event of a compliance audit of your facility's documentation.
5. If you have further questions or concerns after reviewing this notice, please call customer service at 574-371-3071 between 8:00 am and 5:00pm EST, Monday through Friday. Calls received outside of call center operating hours will receive a voicemail prompt or be transferred to an on-call representative in the event of an emergency. Alternatively, your questions may be emailed to CorporateQuality.PostMarket@zimmerbiomet.com.

Other Information

This medical device recall was reported to the U.S. Food and Drug Administration and will be reported to other Competent Authorities, Notified Bodies, and Regulatory Authorities as required.

- Med Watch Reporting: Adverse reactions or quality problems experienced with the use of this product may be reported to the FDA's Med Watch Adverse Event Reporting program either online, by mail, or by fax.
- Online: www.fda.gov/medwatch/report.htm
- Mail: Use postage paid, pre-addressed form FDA 3500, available at: www.fda.gov/MedWatch/getforms.htm
- Fax: 1-800-FDA-0178

Under 21 CFR 803, manufacturers are also required to report any serious injuries where a product has contributed or may have contributed to the event. Please keep Zimmer Biomet informed of any adverse events associated with this product or any other Zimmer Biomet product by emailing product.experience@zimmerbiomet.com.

The undersigned confirms that this notice has been delivered to the appropriate Regulatory Agencies.

Thank you for your assistance. We regret any inconvenience caused by this recall.

Sincerely,

A black rectangular box redacting the signature of the undersigned.

Post Market Surveillance & Regulatory Compliance Director



ATTACHMENT 1

Certificate of Acknowledgement

IMMEDIATE RESPONSE REQUIRED – TIME SENSITIVE ACTION NEEDED

Affected Product: Juggerknot Long Flex Drill Bit with Sleeve Nitinol Stainless Steel

Field Action Reference: ZFA 2018-00252

Do you have affected product in your facility?

- ☐ **Yes**, we currently have one or more affected items in our facility.
- ☐ **No**, we currently have no affected items in our facility.

By signing below, I acknowledge that the required actions have been taken in accordance with this recall notice.

Printed Name: _____ **Signature:** _____

Title: _____ **Telephone:** () _____ - _____ **Date:** ____/____/____

Facility Name: _____

Facility Address: _____

City: _____ **State:** _____ **ZIP:** _____

Note: This form must be returned to Zimmer Biomet before this action is closed for your account. It is important that you complete this form and email a copy to CorporateQuality.PostMarket@zimmerbiomet.com or fax to 574-372-4265.