

Date: 3/4/2016

Field Safety Corrective Action – Product Recall

Product: NK535 – batch no. 52191195
Product description: BIOLOX DELTA REVISION HEAD 12/14 32MM S

Due to feedback from two clinics, it appears that it is possible for size XL adapter sleeves to be in packages of size S Biolox Delta revision heads (see figure 1). This mistake occurs only in individual instances and is restricted to batch 52191195. The batch in question was recalled due to the incorrectly allocated adapter sleeves.




Figure 1: Biolox Delta revision head with adapter sleeves (on the left, correctly packaged sleeves in size S, on the right the incorrectly packaged sleeves in size XL)

Vorsitzender des Aufsichtsrates:
Prof. Dr. h.c. Ludwig Georg Braun

Vorstand:
Prof. Dr. Hanns-Peter Knaebel
(Vorsitzender)
Dr. Dirk Freund
Dr. Joachim Schulz

Sitz der Gesellschaft: Tuttlingen
Reg. Gericht: Stuttgart HRB 726261
USt. Id.-Nr. DE812160059
WEEE-Reg.-Nr. DE 65109852

Bankverbindungen:
Deutsche Bank AG Tuttlingen
BLZ 653 700 75 Konto 21 22 000 00
IBAN DE44 6537 0075 0212 2000 00
SWIFT / BIC DEUTDE33
Baden-Württembergische Bank
BLZ 600 501 01 Konto 487 1905
IBAN DE31 6005 0101 0004 8719 05
SWIFT / BIC SOLADE33

Hausanschrift:
Aesculap AG
Am Aesculap-Platz
78532 Tuttlingen
Deutschland

Page 2 to the letter of March 4, 2016

Since mounting the larger adapter sleeves would result in a length difference of 11.5 m, the danger of the mistake not being noticed during operation is estimated as being extremely small. Therefore, we do not assume there is increased patient risk. Patients already treated using the product are also not at increased risk, since a successful implantation requires that the length of the sleeve is correct for the treatment.

Using the attached return form, please send these articles to:

Aesculap AG
Building 12 – Department QMV / Product Recall
Attn. Jana Zibulski
Am Aesculap-Platz
DE – 78532 Tuttlingen

You will receive the choice of either credit or a replacement delivery. Please fill out the attached return form.

If you have any questions concerning this Field Safety Corrective Action, please contact our product management or our local marketing for orthopaedics/traumatology:

Mr. Thomas Güttler

Tel: + 49 7461 95-2672
Fax: + 49 7461 95-2522
E-Mail: thomas.guettler@aesculap.de

Mr. Dominik Kühne

Tel.: + 49 7461 95-2653
Fax: + 49 7461 95-2131
E-Mail: dominik.kuehne@aesculap.de

Please notify all users of the above product and any other persons who need to be informed within your organisation about this *FSCA*. If you have provided the products to third parties, please forward a copy of this information to them or inform the contact person named below. Please retain this information until the action is closed. The Food and Drug Administration has received a copy of this *FSCA*.

...

Page 3 to the letter of March 4, 2016

Please accept our apologies for the inconvenience caused by this action.

Kind regards,

Aesculap AG

FEEDBACK FORM / FSCA
BIOLOX DELTA REVISION HEAD 12/14 32MM S

Please return this form by fax or e-mail to:

Jana Zibulski / Department QMV

Fax +49 7461-95 1555

vigilance_aag.de@aesculap.de

We confirm having received the product recall of articles NK535. We will implement the recommended measures:

- We do not have any of the affected products in use.
- We send back the affected products from our establishment and wish to receive a:
 - credit note.
 - replacement delivery.

HOSPITAL _____

CITY _____

NAME _____

DEPARTMENT _____

PHONE _____

SIGNATURE

DATE