

07 Nov 2016



URGENT FIELD SAFETY NOTICE: RA2016-169

FSCA Identifier:	Product Field Action RA 2016-169
Type of Action:	Field Safety Corrective Action
Description:	sterile packaging of Guide Wire(s)
Legal Manufacturer	Stryker Trauma GmbH, Prof.-Kuentscher-Strasse 1-5, 24232 Schoenkirchen/Kiel, Germany
License holder	Stryker Singapore Pte Ltd, 108 Pasir Panjang Road #03-04 Golden Agri Plaza Singapore 118535
Catalogue number	Product name
12106450S	GAM Kirschner Wire
18060050S	T2 K-Wire
18063030S	T2 K-Wire Recon

Lot #s: 51 specific lots as per attached Affected Product List.

Dear Customer,

Please find attached details of a voluntary Field Safety Corrective Action that has been initiated by Stryker Trauma GmbH, Division Trauma and Extremities for sterile packaging of K-Wire(s). It was found through review of packaging that the seal integrity of the pouch may be compromised. More specifically, there is a potential that the sterile pouch is not sealed at one end due to a manufacturing error.

Stryker Trauma GmbH, Division Trauma and Extremities is recalling all unconsumed, non-expired lots of above listed article numbers. Given high turnover for this product and the frequency with which it had been on backorder it is not expected that a significant quantity of units subject to this notice remain in the field. No injury or harm has been reported for this event.



Packaging example – sterile pouch inside (plastic) clear tube



Example of chevron seal - potentially not present

Potential Hazards

A missing seal could potentially lead to unsterile product.

Mitigating Factors

The nonconformance is obvious to the user.

Surgical guidelines outline inspection of the sterile barrier (seal) for sterile packed medical devices prior to use.

The pouch itself shows a note: "Contents sterile unless this package has been damaged or opened."

The secondary packaging is a (plastic) clear tube with silicone caps at both ends. While not validated as a sterile barrier, it does provide additional protection to the enclosed pouch package configuration.

Furthermore, it should also be noted that it is standard practice for surgeons to administer antibiotics peri-operatively in order to reduce the risk of potential infection.

Our records indicate that you have received at least one of the subject devices and you are therefore affected by this action.

We request that you read this notice carefully and complete the following actions:

1. Immediately check your internal inventory and quarantine all subject devices pending to return to Stryker.
2. Circulate this Field Safety Notice internally to all interested/affected parties.
3. Maintain awareness of this notice internally until all required actions have been completed within your facility
4. Inform Stryker if any of the subject devices have been distributed to other organisations.
 - a) *Please provide contact details so that Stryker can inform the recipients appropriately*
 - b) *If you are a Distributor, note that you are responsible for notifying your affected customers.*
5. Please inform Stryker of any adverse events concerning the use of the subject devices.
Please comply with any local regulations concerning the notification of adverse events to your National or local Competent Authorities.
6. Complete the attached customer response form. It may be that you no longer have any physical inventory on site. Completing this form will allow us to update our records and will also negate the need for us to send any further unnecessary communications on this matter. Therefore please complete even if you no longer have any of the subject devices in your physical inventory.

7. Return the completed form to your nominated Stryker Representative (indicated below) for this PFA. On receipt of the form, a Stryker Representative will contact you to organize any applicable ongoing actions.

If returning the product would adversely impact your ability to provide necessary medical care to patients, you can consider re-sterilizing the product per Sterilization instructions contained in the Instruction for Use.

We request that you respond to this notice within 7 calendar days from the date of receipt. The target date for completion of this action is ----- and your timely response will enable us to ensure that we meet this target.

Please report any adverse events or product quality problems associated with this device to Stryker. Healthcare Professionals may also report any suspected adverse events associated with these devices to the Vigilance Branch, Health Products Regulation Group, HAS at Tel: 6866 3538, Fax: 6478 9069, or report online at www.hsa.gov.sg/ae_online.

On behalf of Stryker we thank you sincerely for your help and support in completing this action within the target date and regret any inconvenience that may be cause. We would like to reassure you that Stryker is committed to ensuring that only conforming devices, meeting our high internal quality standards, remaining on the market.

Yours Sincerely,



Sara Jato
QA Specialist
Stryker ASEAN
sara.jato@stryker.com
ASEAN.PMS@stryker.com
Tel. +84 (0)8 3827 5399 Ext.12

RA2016-169 Affected Product and Lot Codes

Manufacturer Part Number	Manufacturer Part Name	Lot Numbers			
12106450S	GAM Kirschner Wire	K0800BB	K081727	K0911F4	K09D564
		K0800BC	K082C8B	K09379B	K09F026
		K0800BD	K084F89	K096A26	K0A1EF8
		K0800BE	K084F98	K096A2A	K0A1EFA
		K0800BF	K08683D	K096A2C	K0A1EFB
		K0800C0	K08683E	K098213	K0A38FB
		K081720	K086841	K098215	K0A63AB
		K081721	K08820B	K099AA0	K0A63AC
		K081722	K089AF5	K09AD33	K0A7BC1
		K081723	K08E1E1	K09BA4F	
18060050S	T2 K-Wire	K084FBF	K0937A7	K09BA53	K09BA55
		K0920A4	K0937C4	K09BA54	K09BA56
		K0A1EFD			
18063030S	T2 K-Wire Recon	K086847	K08E1E9	K099AA8	

FIELD SAFETY CORRECTIVE ACTION ACKNOWLEDGMENT FORM

FSCA Identifier: Product Field Action RA 2016-169

Type of Action: Field Safety Corrective Action

Description: sterile packaging of K-Wire(s)

Legal Manufacturer Stryker Trauma GmbH, Prof.-Kuentscher-Strasse 1-5,
24232 Schoenkirchen/Kiel, Germany

Catalogue #s: Refer to the attached list

Lot #s: Refer to the attached list

I acknowledge receipt of the Field Safety Notice for RA2016-169 and can confirm that:

We have not located any of these devices in our inventory: <i>(please delete if not applicable)</i>				
We have located the following devices:				
Product description	Product Reference	Lot Number	Qty	Qty Quarantined
We have further distributed subject devices to the following organisations:				
Facility Name				
Facility Address				
Form completed by:				

Contact Name	_____	Contact Facility	_____
Contact address	_____	Contact Position	_____
	_____	Contact Tel No	_____
	_____	Contact Fax No	_____
	_____	Contact e-mail	_____

PLEASE COMPLETE AND EMAIL TO ASEAN.PMS@STRYKER.COM.