

**URGENT MEDICAL DEVICE RECALL NOTIFICATION**  
**Dri-Lok Threaded Cannula**

December 27, 2017

Account Number:

**Attn:**

The purpose of this notification is to advise you that Stryker Endoscopy is conducting a voluntary recall.

**Part Number:** 3910-090-800, ASSY, PKG, 8.0MM X 90MM DRI-LOK THREADED CANNULA

**Lot Number:** 14239AG2

**Product Description:**

The Dri-Lok Cannulas are single-use devices which facilitate instrument insertion and removal during arthroscopic surgery.

**Reason for the Voluntary Recall:**

Stryker inadvertently shipped expired units to two customers. Therefore, we are initiating this voluntary recall to recover the two boxes that were shipped.

**Risk to Health:**

The risk of using a Dri-Lok cannula beyond its expiration date is that the product may no longer be sterile. As a result of potential ineffective sterilization, an infection may occur. To date, there have been no reports of any adverse events or serious injuries.

**Actions to be taken by the Customer/User:**

1. Inform individuals within your organization who need to be aware of this device recall.
2. Check all stock areas and/or operating room storage to determine if any devices from the affected product list are at your facility. **Response is required.**
3. Quarantine and discontinue use of the recalled devices.
4. If no product is found, please complete Attachment A on the back side of this letter.
5. If affected product is found, segregate the product and call Stryker customer service at 1-800-624-4422 (Option 3) or email [endocustomersupport@stryker.com](mailto:endocustomersupport@stryker.com) to arrange for product return and issuance of credit or replacement (upon availability).

Please forward a copy of this letter to any other personnel within your facility that you deem appropriate.

We appreciate your cooperation and we recognize the inconvenience this may cause your facility. Thank you for your support on this important matter.

Sincerely,



Kimberly Lynch

**Attachment A**



**By completing this form, you are acknowledging none of the affected product below at your account. All sections below must be completed in order for acknowledgement to be accepted.**

**Part Number, Description:** 3910-090-800, ASSY, PKG, 8.0MM X 90MM DRI-LOK THREADED CANNULA  
**Lot Number:** 14239AG2

Return the completed Business Reply Form to Stryker Endoscopy via fax (408-855-6314) or

Account Name	
Account Number	
Account Address	
Name & Title	
Signature	
Date	
Phone or Email Address	

email ([endorecall@stryker.com](mailto:endorecall@stryker.com)). Please send any questions to [endorecall@stryker.com](mailto:endorecall@stryker.com)