

Valeant Pharmaceuticals International, Inc.
 400 Somerset Corporate Boulevard, Bridgewater, NJ 08807
 Tel: 908-927-1400, Fax: 908-927-1856
www.valeant.com

URGENT VOLUNTARY PRODUCT RECALL AND FIELD CORRECTION TO LASEREDGE 2.0mm X 2.2mm TRAPEZOIDAL ANGLED, E7601 Lot MAPC520

Friday, 19 February 2016

Re: Notice of Dull Blades for LASEREDGE 2.0mm X 2.2mm TRAPEZOIDAL ANGLED 6/Box SKU# E7601 Lot MAPC520

Product Details: 2.0mm X 2.2mm Trapezoidal Angled Knife with handle (SKU# E7601 - 6 per box) manufactured by Angiotech (Surgical Specialties)


Product Name: LASEREDGE 2.0mm X 2.2mm Trapezoidal Angled 6/Box SKU#E7601

Lot numbers of affected product shipped to you: MAPC520

Please Note: This recall is limited to E7601 Lot# MAPC520 only and does not affect other lots shipped to your facility

Dear Valued Customer

This is to inform you of a medical device voluntary product recall and field correction involving LASEREDGE 2.0mm X 2.2mm TRAPAZOID ANGLED 6/Box, SKU#E7601 Lot MAPC520. The product label is shown below to assist you in identifying the product within your facility.



Valeant Pharmaceuticals International wants to ensure that our products provide customer satisfaction and ease of use. It has come to our attention that some boxes of LASEREDGE 2.0mm X 2.2mm Trapezoidal Angled Knives 6/Box, SKU# E7601 Lot MAPC520 have not been as sharp as

previous lots of this product. Please review carefully the notes outlined in this letter regarding your LASEREDGE 2.0mm X 2.2mm Trapezoidal Angled Knives 6/Box, SKU# E7601 Lot MAPC520.

This action represents a voluntary product recall and we have notified the appropriate authority of this voluntary recall.

Please take the following steps to ensure the recalled product is handled appropriately:

1. **Quarantine the product:** According to our records, your facility may have a supply of LASEREDGE 2.0mm X 2.2mm Trapezoidal Angled Knives 6/Box, SKU# E7601 Lot MAPC520. Please quarantine any unused boxes (full and partial) of the above product that you have in your facility. All unused boxes of LASEREDGE 2.0mm X 2.2mm Trapezoid Angled Knives 6/Box, SKU# E7601 Lot MAPC520 shall be returned to Bausch & Lomb UK Limited using the Freepost label enclosed.

2. **To obtain replacement product**


To obtain replacement product free of charge please acknowledge receipt of this notification by completing and returning the attached Recall and Field Correction Acknowledgement Form, with the date of this letter's receipt noted.

When completing this form be certain to:

- Count the number of product packages in quarantine and detail this on the form.
- Indicate your choice for receiving replacement product
- Email the form to uksurgorders@bausch.com or fax to 020 8781 0001

We sincerely apologise for any inconvenience and assure you that we are working diligently to resolve this issue in a timely manner. Should you have any questions or concerns, please feel free to contact customer services on 020 8781 0000 or uksurgorders@bausch.com

Yours faithfully



Director Food Supplement
Cosmetics & Medical Devices
Valeant EMEA

Recall and Field Correction Acknowledgement Form

**Product Details: LASEREDGE 2.0mm X 2.2mm Trapezoid Angled Knives 6/Box
SKU#E7601 Lot MAPC520**

This is to acknowledge receipt of the above referenced recall and field correction notification dated Friday, 19 February 2016

Please confirm inventory levels of the affected cannulas with 4-digit lot numbers at your facility:

Product	Lot #	Qty Shipped	Qty Used	Qty in Inventory	Responsible Person's Initials
E7601	MAPC520				

If you have inventory of affected E7601, Lot MAPC520 at your facility and would like to receive replacement product, please check the box below :

I hereby certify that I have returned the existing inventory of E7601, Lot MAPC520 in our facility and ask that replacement product equal to the amount of quarantined product shown above, referencing PO# _____, be shipped to our facility.

Date		Name (Print)	
Account Number	«ACCOUNTNO»	Signature	
Facility Name	«ACCOUNT»	Telephone Number	

Please complete, sign and return this Form to:

Email: uksurgorders@bausch.com

Fax: 020 8781 0000