

March 19, 2018

MEDICAL DEVICE RECALL- 5 mm Apple-Hunt Secondary Cannula


Dear Valued Customer:

CooperSurgical is recalling a single lot of the 5 mm Apple-Hunt Secondary Cannula (P/N: 900-840 Lot: 223333). The products were identified to have been packaged with the wrong Trocar length type. This discrepancy does not meet the product release specifications. Our records indicate that one or more of the affected products list below may have been distributed to your facility.


PART NUMBER	PART DESCRIPTION	LOT NUMBER
900-840	5 mm Apple-Hunt Secondary Cannula/Pyramidal Tip Trocar - extended length (5/Box)	223333

The Apple-Hunt Secondary Cannulas are indicated for use in laparoscopic surgical procedures to provide a pathway for the introduction of endoscopic surgical instruments and allow for maintenance of pneumoperitoneum. The product package contains 5 individually Tyvek-sealed clear plastic trays. The device lot number and expiration date is located on the front Tyvek lid as well as on the side of the external packaging.

TRAY



BOX



The event was discovered internally during a complaint investigation. No adverse events or injuries have been reported to date in connection with the use of the 5 mm Apple-Hunt Secondary Cannula. We apologize for any inconvenience this may have caused.

Please discontinue use of any of the lots listed in the table above and complete the attached **Acknowledgement and Receipt Form** to schedule refund or exchange. If you have supplied or transferred any potentially affected product to another facility or organization, please let them know of the recall by providing a copy of this letter. If you have any further questions, please feel free to contact CooperSurgical at recall@coopersurgical.com.

Sincerely,

[Redacted Signature]

Nana Banarji
Manager Clinical Product Surveillance