

Smith & Nephew, Inc.
1450 Brooks Road
Memphis, TN 38116
USA

1-901-396-2121
1-800-821-5700
www.smith-nephew.com



Urgent Medical Device Recall Notice
R-2017-22

September 14, 2017

<Insert Address>

This letter is to inform you that Smith & Nephew, Inc. has initiated a field action to voluntarily remove a single lot of LEGION SCREW-ON WEDGE SIZE 4 due to a manufacturing error. The Legion Screw-on Wedge Size 4 contained 5MM screws instead of the required 10MM screw.

Please see product details below:

Product Number	Description	Batch Number
71421733	LEGION SCW LWDG S4 10D X 5P	14GSM0440A

Shipment Date: December 23, 2014 through April 25, 2017

Potential Risk with Use of the Product

In the event the affected product is presented for use and a backup is not available the surgeon would have the option to make additional bone cuts to upsize the L-wedge. This could potentially lead to extended surgery time and require additional bone sacrifice.

Required Actions:

- Please follow the instructions on the attached Response Form.

Enclosure: Response Form



Urgent Medical Device Recall Notice

R-2017-22

Response Form

September 14, 2017

<Insert Address>

PLEASE COMPLETE ALL ITEMS AND RETURN WITHIN 5 DAYS OF RECEIPT

Required Actions:

1. Please inspect your inventory and locate any unused devices from the listed product and batch numbers on the first page of this Field Action Notification, and quarantine them immediately.
 - a. If you are a distributor, you must notify your customers of the field action and ensure that these actions are carried out.
2. If you have no product to return, please put an X in the appropriate location below.
3. If you have product to return, please list the batch numbers and quantities of each batch that you are returning in the appropriate boxes below.
4. Complete the remainders of the form sign and send to FieldActions@smith-nephew.com or fax to 901-566-7975.

Please Note – even if you have no product to return, this form must be completed, signed and returned.

5. Once the form is received by Smith & Nephew, you will be sent a Return Authorization (RA) number.

If you have any questions or concerns regarding this recall please contact FieldActions@smith-nephew.com.

No Product to Be Returned

☐

Product Part Number	Batch Number (List Specific Batch #'s to be Returned)	Quantity of Units to be Returned

We hereby confirm that we are aware of this Medical Device Field Action and it has been communicated within our organization.

Printed Name (required): _____ Title: _____

Signature (required): _____ Date (required): ____/____/____

Email: _____ Telephone: (____) _____ - _____

S&N Account Number: _____ RA Number (S&N use only): _____

Name of Organization(s) Covered by Response: _____

Return affected product to: Smith & Nephew | Attn: Global Field Actions | Building G, 1450 Brooks Rd. East | Memphis, TN 38116