Leica Biosystems Newcastle Ltd Balliol Business Park West Benton Lane Newcastle upon Tyne NE12 8EW United Kingdom

Tel. +44 (0)191 215 0567 Fax. +44 (0)191 215 1152



Urgent Field Safety Notice

Product Name: CD10-270-CE-S and CD10-270-CE Novocastra™ Lyophilized Mouse Monoclonal

Antibody CD10 Reagents

Lot(s) #: 6033086, 6032390, 6033622, 6032742

Reason:

Instability of Lots

Date:

03 Sep 2015

Attention:

Pathology Department / Dealer

Details on affected devices:

This Urgent Field Safety Notice is applicable to the following products:

Product Code	Lot Numbers	
CD10-270-CE-S	6033086, 6032390, 6033622	
CD10-270-CE	6032742	

Description of the problem:

Leica Biosystems recently became aware that the Novocastra[™] Lyophilized Mouse Monoclonal Antibody CD10 reagent lots listed above do not function as intended up to the expiry date on the product labelling.

This was detected through stability monitoring carried out by Leica Biosystems. There is a link between the age of the product and staining intensity.

This issue is limited to this products and the lot numbers listed above.

As a result of this situation, Leica Biosystems is issuing this Urgent Field Safety Notice.

ACTIONS REQUIRED

Do not use or continue to use the reagent lots listed above, as they have the potential to result in inadequate tissue staining and lead to weak or false negative pathology results. It is requested that customers appropriately destroy any unused or partially used affected lots of the reagent, and indicate on the attached Urgent Field Safety Notice Acknowledgement Form that this action has been undertaken.

As indicated in the Instructions for Use, the clinical interpretation of any staining or its absence should be complemented by morphological studies using proper controls and should be evaluated within the context of the patient's clinical history and other diagnostic tests by a qualified pathologist. Consult with your laboratory director and the physician to determine whether a review of past results obtained from the affected lots is clinically warranted.

Please contact your local Leica representative if there are any questions or concerns, or alternatively, please contact us at the following email address;

CD10-270@Leicabiosystems.com

Leica Biosystems Newcastle Ltd Balliol Business Park West Benton Lane Newcastle upon Tyne NE12 8EW United Kingdom

Tel. +44 (0)191 215 0567 Fax. +44 (0)191 215 1152



Transmission of this Urgent Field Safety Notice:

Kindly pass this notice primarily to the end users where the products have been sold and to all those within your organisation who need to be aware of this issue.

Please confirm receipt of this notice as soon as possible by signing and dating the attached Urgent Field Safety Notice Acknowledgement Form and fax it back to the following number, to receive your alternative replacement for applicable orders.

+44 191-215-1152

Alternatively, scan and return using

CD10-270@Leicabiosystems.com

An alternative replacement for applicable orders will be processed by Leica Biosystems, upon receipt of completed acknowledgement form.

Your cooperation in this matter is greatly appreciated and we sincerely apologise for any inconvenience this may have caused.

Regards,

Contact reference person:

Peter Lloyd Leica Biosystems Newcastle Ltd Balliol Business Park West Benton Lane Newcastle upon Tyne NE12 8EW United Kingdom

Telephone: +44 191 215 0567 Facsimile: +44 191 215 1152

The undersigned confirms that the relevant Competent Authorities are aware.

Signature



Leica Biosystems Newcastle Ltd Balliol Business Park West Benton Lane Newcastle upon Tyne NE12 8EW United Kingdom

Tel. +44 (0)191 215 0567 Fax. +44 (0)191 215 1152



URGENT FIELD SAFETY NOTICE ACKNOWLEDGEMENT FORM

The following lot numbers are affected:

Product Code	Lot Numbers
CD10-270-CE-S	6033086, 6032390, 6033622
CD10-270-CE	6032742

I hereby acknowledge receipt of the Leica Biosystems Field Safety Correction Notification, and that affected lot(s) have been destroyed.

An alternative replacement for applicable orders will be processed by Leica Biosystems, upon receipt of completed acknowledgement form.

Contact Person (Please Print)	Facility (Please Print)
	Address (Please Print)
Signature	 Date

Please fax or email this form to the following:

+44 191-215-1152

CD10-270@Leicabiosystems.com