

Thermo Fisher Scientific
Microbiology
12076 Santa Fe Trail Drive, P.O. Box 14428
Lenexa, KS 66215
800.255.6730

877.428.1924 fax www.thermoscientific.com



URGENT MEDICAL DEVICE FIELD ACTION RECALL ATTN LABORATORY: RECALL INFORMATION – ACTION REQUIRED

08/13/2018

«Ship_To_Customer_Address_1»

«Ship_To_Customer_Address_2»

«Ship_To_Customer_Address_3»

 ${\it ``Ship_To_Customer_Address_4"}$

URGENT: DELIVER TO MICRO LAB

Product Number	Product Name	Lot Number	
R05152	Thio med w/ dex, hem, vit k (7ml)100/pk	258192	
R064700		254751	
	Thio med w/o ind w/dex (9ml) 100/pk	259508	
		272498	
R064702	Thio med w/o ind, w/ dex (9ml) 20/pk	274247	
	Thio med w/ dex, hem, vit k (9ml)100/pk	254749	272497
R064720		251145	272015
		258183	258626
		271479	258625
R064730	Thio med w/ dex, hem, vit k (5ml) 100/pk	254755	272020
		260509	258181
		273564	258627
		271476	258629
		271665	259507
		251146	272499
R064732	This mad w/ day hamin with (5ml)20/pla	271667	
	Thio med w/ dex, hemin, vit k (5ml)20/pk	258182	
R07176	Thio med w/o ind w/dex (10ml) 100/pk	259521	
	Thio med w/ dex, hem, vit k(10ml) 100/pk	259526	258906
R07180		258903	
		258904	
		258905	
R07182	This mad w/ day have with (10ml) 20/nls	258907	260413
KU/182	Thio med w/ dex, hem, vit k (10mL) 20/pk	258908	260519

REASON FOR FIELD ACTION:

Customers are to be advised of the following:

An internal technical investigation has confirmed that **the products and lot numbers list above**, have intermittently failed performance testing during stability studies with various anaerobic microorganisms.



RISK TO HEALTH:

Thio media is recommended for use in qualitative procedures as a general purpose media for the cultivation of aerobes and anaerobes.

We believe the risk is low based on the following rationale:

The immediate health consequence may be a delay in the recovery of anaerobic bacteria associated with an infectious process. The risk is mitigated by the empiric treatment for anaerobic infections prior to bacteria isolation in patients already suspected of having an anaerobic infection. This is mainly due to the longer incubation time for anaerobes, which can be up to several days. This empiric treatment reduces the probability that patients with suspected anaerobic infections are at greater risk of impaired clinical response.

Furthermore, no complaints nor reports of illnesses or injuries have been received.

ACTIONS TO BE TAKEN BY THE CUSTOMER:

Our records indicate that you have received the above products.

Please notify any personnel who need to be aware of the potential for failure of performance for this product. Accordingly, in keeping with our Quality Policy, we request that you inspect your stock and destroy any remaining inventory of the lots listed above and contact Customer Services or your local distributor regarding replacement. You should review results and laboratory reporting associated with use of the listed lots and consider retesting and/or seek appropriate expert advice for further action.

The US Food and Drug Administration (FDA) has been informed of this Field Action Recall.

Please return the attached Medical Device Recall Return Response form acknowledging receipt of the notice and disposal of the affected product to help with our FDA reporting obligation.

This notice needs to be passed on to all who need to be aware within your organization or to any organization where the potentially affected products have been transferred. If you have any questions, please contact our Technical Services Department at 800-255-6730 (US) or 913-888-0939 (International).

Thermo Fisher SCIENTIFIC

PRODUCT AND DISTRIBUTION INFORMATION:

Product Name	Product Number	Lot Number	Distribution Date (MMDDYYYY)	Expiration Date (MMDDYYYY)	Distributed Total Quantity
Thio med w/ dex, hem, vit k (7ml) 100/pk	R05152	258192	12/26/2017 to 02/14/2018	06/28/2018	48
Thio med w/o ind w/dex (9ml) 100/pk		254751	12/18/2017 to 02/06/2018	08/27/2018	104
	R064700	259508	01/02/2018 to 02/19/2018	09/04/2018	102
		272498	01/30/2018 to 03/06/2018	10/04/2018	19
Thio med w/o ind, w/ dex (9ml) 20/pk	R064702	274247	02/20/2018 to 06/07/2018	10/09/2018	150
-		254749	12/04/2017 to 12/13/2017	08/17/2018	94
		251145	11/28/2017 to 12/04/2017	08/27/2018	77
		258183	12/11/2017 to 01/15/2018	08/30/2018	98
Thio med w/ dex, hem,	D064720	271479	01/22/2018 to 02/13/2018	08/31/2018	53
vit k (9ml) 100/pk	R064720	272497	01/30/2018 to 02/19/2018	08/31/2018	39
		272015	01/24/2018 to 02/19/2018	10/02/2018	42
		258626	12/07/2017	10/03/2018	53
		258625	12/11/2017 to 01/15/2018	10/04/2018	38
		254755	12/14/2017 to 01/04/2018	08/17/2018	67
		260509	12/12/2017 to 01/04/2018	08/27/2018	110
		273564	01/24/2018 to 02/26/2018	08/30/2018	111
		271476	01/10/2018 to 02/05/2018	08/31/2018	86
Thio med w/ dex, hem, vit k (5ml) 100/pk	R064730	271665	01/18/2018 to 02/06/2018	08/31/2018	68
		251146	11/27/2017 to 12/11/2017	08/31/2018	58
		272020	01/09/2018 to 01/17/2018	08/31/2018	151
		258181	12/11/2017 to 12/21/2017	09/04/2018	59
		258627	12/11/2017 to 02/05/2018	09/04/2018	59
		258629	12/07/2017 to 12/26/2017	09/06/2018	82
		259507	12/12/2017	10/02/2018	75
		272499	01/29/2018 to 02/19/2018	10/08/2018	85
Thio med w/ dex, hemin,		271667	01/12/2018 to 03/19/2018	08/30/2018	343
vit k (5ml) 20/pk	R064732	258182	12/11/2017 to 01/15/2018	10/02/2018	343
Thio med w/o ind w/dex (10ml) 100/pk	R07176	259521	01/17/2018 to 02/21/2018	09/04/2018	9
(10mm) 100/pm		259526	12/27/2017 to 02/20/2018	09/03/2018	146
		258903	12/18/2017 to 01/29/2018	09/03/2018	22
Thio med w/ dex, hem,	R07180	258904	12/18/2017 to 01/17/2018	09/03/2018	23
vit k (10ml) 100/pk	K07100	258905	12/11/2017 to 01/03/2018	09/03/2018	21
		258906	12/11/2017 to 01/03/2018 12/12/2017 to 01/10/2018	09/04/2018	22
		258907	12/12/2017 to 01/10/2018 12/12/2017 to 03/12/2018	09/03/2018	106
Thio med w/ dex, hem,		258908	12/11/2017 to 04/03/2018	09/03/2018	122
vit k (10mL) 20/pk	R07182	260413	01/29/2018 to 04/03/2018	09/05/2018	43
vit k (10IIIL) 20/pk		260519	02/05/2018 to 02/27/2018	09/06/2018	34



TYPE OF ACTION TO BE TAKEN BY THE MANUFACTURER:

Investigation into the root cause is currently on-going. Corrective actions will be implemented upon completion of the investigation.

We appreciate your immediate attention to this field correction. We apologize for any inconvenience this may have caused and appreciate your understanding as we take action to ensure customer safety and satisfaction.

If you have any questions, please contact our Technical Services Department at 800.255.6730 (US) or 913.888.0939 (International).

Sincerely,

Gary Klaassen

Director Quality Assurance & Regulatory Affairs



MEDICAL DEVICE RECALL/FIELD SAFETY NOTICE RETURN RESPONSE

Acknowledgment Form *Response Required

CUSTOMER INFORMATIO

Account Number: «Ship_To_Account» «Ship_To_Customer_Address_1» «Ship_To_Customer_Address_2» «Ship_To_Customer_Address_3»
«Ship_To_Customer_Address_4»
I have read and understand the attached customer letter and field action instructions: (initials)
I understand that this applies to all inventory of the affected product that I have received: (initials)
Has the recalled product caused and/or contributed to any deaths or serious injuries? Yes No If yes, please explain:

AFFECTED PRODUCT REFERENCE NUMBER:

See below for products listed below that were impacted by this notice:

Product Number	Lot Number	Units Received	Units Discarded
R05152	258192		
	254751		
R064700	259508		
	272498		
R064702	274247		
	254749		
	251145		
R064720	258183		
	271479		
	272497		
	272015		
	258626		
	258625		
R064730	254755		
	260509		
	273564		
	271476		
	271665		

Product Number	Lot Number	Units Received	Units
Number		Received	Discarded
	251146		
	272020		
D064720	258181		
R064730	258627		
	258629		
	259507		
	272499		
R064732	271667		
	258182		
R07176	259521		
	259526		
	258903		
D07100	258904		
R07180	258905		
	258906		
R07182	258907		
	258908		
	260413		
	260519		

RESPONSE BELOW.

Use additional sheet(s) if necessary.



#