

## Customer Notification Letter

Please forward this information to all relevant users and biomedical staff in your facility.

### Subject: Symbols for manufacturing and expiry date switched on disposable CO2 absorbers

#### Products affected:

Part #	Description	Lot# / SN
6677845	Disposable absorbers, 12 pcs	1810 and 1811

Dear Customer,


cc to the Chairman Medical Board and relevant Head-of-Department


On some of our disposable CO2 absorbers shipped from our warehouse between January 28 to May 22 2019, the expiry date and the manufacturing date have switched places stating that the absorber expired before it was produced (Refer to picture below). Apart from this, there is nothing wrong with the absorber and they are safe to use.

Nevertheless, if you have such disposable absorbers in your stock (see lot numbers above in the table) and wish to replace them free of charge, kindly contact your local Getinge representative.

Your local Getinge representative will also be able to answer any questions you may have related to this matter.

Yours sincerely,

  
Andras Sarkany  
Product Manager  
Manufacturer

  
Pernilla Sandberg  
Head of Quality  
Manufacturer

