



Chemical Metrology Division
Applied Sciences Group
Health Sciences Authority
1 Science Park Road, #01-05/06,
The Capricorn, Singapore Science Park II,
Singapore 117528
Tel: 65 6775 1605 Fax: 65 6775 1398
Website: www.hsa.gov.sg
Email: HSA_CML@hsa.gov.sg



Ref. No.: CML-HRM-2018A/02
Date of Issue: 20 Feb 2025

Certificate of Analysis

CERTIFIED REFERENCE MATERIAL HRM-2018A

Elements in Lipstick Material

ARMN Code

ARM M001

Batch Number

STY-0137-001

Description

The certified reference material (CRM) was produced by spiking lipstick material with solutions of $\text{Na}_2\text{HAsO}_4 \cdot 7\text{H}_2\text{O}$, $\text{Cd}(\text{NO}_3)_2 \cdot 4\text{H}_2\text{O}$ and $\text{Pb}(\text{NO}_3)_2$. The material (10 g) was bottled in a brown glass bottle and sealed with a screw cap.

The reference material was produced with reference to the requirements set out in ISO/IEC 17025:2017 [1], ISO 17034:2016 [2] and ISO Guide 35:2017 [3].

Certified Mass Fraction Values

A certified value is a value for which a laboratory has the highest confidence in its accuracy, in that all known or suspected sources of biases have been investigated and accounted for. The certified mass fraction values for the three elements in the CRM are listed below. The certified mass fraction values for cadmium (Cd) and lead (Pb) were determined by inductively coupled plasma mass spectrometer using isotope dilution mass spectrometry (ICP-IDMS) [4]. The certified mass fraction value for arsenic (As) was determined by inductively coupled plasma high resolution mass spectrometer (ICP-HR-MS) using standard addition method [5].

Analyte	Mass fraction	Unit
Arsenic	6.22 ± 0.42	mg/kg
Cadmium	3.08 ± 0.15	mg/kg
Lead	14.89 ± 1.07	mg/kg

The mass fraction value is expressed as the certified value $\pm$ the expanded uncertainty.

The uncertainty listed with the certified value is an expanded uncertainty about the mean, with

coverage factor 2 (approximately 95% confidence). The certified value has an associated measurement uncertainty attributed to uncertainty contribution from characterisation of the material (u_{char}), uncertainty in the homogeneity of the material (u_{bb}) and uncertainty in the stability of the material (u_{stab}). The u_{char} was evaluated by combining uncertainties from method precision, the concentration of calibration solution, weighing and different ion pairs used (for Cd and Pb), isotope ratios (for Cd and Pb) and the relative atomic mass (for Pb only), in accordance with ISO/IEC Guide 98-3:2008 [6].

Homogeneity

Homogeneity testing on the analytes in the material was performed on at least eleven bottles with two sub-samples taken from each bottle. ICP-MS was employed for the determination of the analytes. The sample size taken for homogeneity testing was about 0.35 g. No significant differences in the between and within-bottle variances were found using one-way ANOVA at 95 % confidence level [3]. Thus, the material was regarded to be sufficiently homogeneous.

Stability

Short-term stability testing on the analytes in the material at 50 °C (maximum allowable transportation temperature) showed that they were stable up to 21 days.

The long-term stability of the analytes at storage temperature (18 °C to 25 °C) was evaluated on ten occasions over a period of up to 22 months after preparation. The results showed that the analytes were stable over the study period.

Validity of Certified Mass Fraction Values

The certified mass fraction values are valid within its measurement uncertainty until **10 Mar 2028**, provided that the CRM is subjected to the same handling and storage conditions as stated in this Certificate of Analysis (COA).

The CRM will be continuously monitored during the validity period to determine if any substantive change to the certified values has occurred. If necessary, its user will be advised or an updated COA will be issued when the property value of the CRM is found to have changed.

Analytical Methods

The certified mass fractions of Cd and Pb in the material were determined by exact-matching ICP-IDMS. Standard reference materials from the National Institute of Standards and Technology (NIST, USA) (SRM 3108 for Cd and SRM 3128 for Pb) were used as calibration standards for IDMS measurements. Enriched isotopes ^{111}Cd and ^{206}Pb from Oak Ridge National Laboratory (USA) were used as the internal standards. The calibration blends were prepared gravimetrically by mixing appropriate amount of calibration standard solutions and internal standard solutions. The sample blends were prepared by spiking appropriate amount of internal standard into the material.

The certified mass fraction of As was determined by ICP-HR-MS using standard addition method. Gallium standard reference material from NIST (SRM 3119a) was added to the sample digest as internal standard. Different amounts of As standard reference material from NIST (SRM 3103a) were then spiked into the sample digest to produce sample blends. Quality control blends were also prepared and analysed concurrently. Both sample and quality control blends were subjected to microwave acid digestion using 5 mL HNO_3 , 0.2 mL HF and 2 mL H_2O_2 to give a clear digest.

Metrological Traceability

The certified mass fraction values are traceable to the International System of Units (SI) through the use of standard reference materials from NIST.

Intended Use

For the validation of methods or as quality controls used to determine the mass fraction of As, Cd and Pb in cosmetic products with high organic content, such as lipstick.

Instruction for Use

After use, the bottle should be re-capped, sealed with Parafilm and stored properly under the recommended storage conditions (please see below). The minimum sample size for each use should be about 0.35 g.

Storage

The material should be stored at 18 °C to 25 °C in its original bottle. Exposure to direct intense light and ultraviolet radiation should be avoided.

Safety Precautions for Users

Treat the material as hazardous substance. Use appropriate work practices when handling to avoid skin or eye contact or ingestion.

Further Information

Please direct all enquiries regarding this CRM to the contact provided in this COA.

References

1. ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories.
2. ISO 17034:2016 General requirements for the competence of reference material producers.
3. ISO Guide 35:2017 Reference materials – Guidance for characterisation and assessment for homogeneity and stability.
4. Sargent, M.; Harrington, C.; Harte, R.; *Guidelines for Achieving High Accuracy in Isotope Dilution Mass Spectrometry*, RSC Publishing, 2002.
5. Abbyad, P.; Tromp, J.; Lam, J.; Salin, E.; *Optimization of the technique of standard additions for inductively coupled plasma mass spectrometry*, J. Anal. At. Spectrom. (2001) 16: 464 – 469.
6. ISO/IEC Guide 98-3:2008 Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995).

Additional Statement and Acknowledgement

This CRM was produced as a part of the ASEAN Reference Material Network (ARMN) joint project. The certified mass fraction values and associated uncertainties stated in this COA were determined by HSA and confirmed by the National Institute of Metrology (Thailand) (NIMT). NIMT also carried out measurements to confirm the short-term and long-term stability of this material.

Certificate Revision Records

Certificate Ref. No.	Date of issue	Reason for issuance
CML-HRM2018A/01	10 Mar 2022	Issuance of first certificate
CML-HRM2018A/02	20 Feb 2025	Extension of expiry date

Note

HSA does not assume any liability with respect to any loss caused by improper use and/or storage of the reference material by the customer.

A handwritten signature in black ink, appearing to read 'Teo Tang Lin', with a horizontal line drawn through the middle of the signature.

Dr Teo Tang Lin
Division Director
Chemical Metrology Laboratory
Chemical Metrology Division